

ELISpot Flex: Human IgG2 (HRP)

3852-2H

The kit is intended for enumeration of cells secreting human IgG2 using the ELISpot assay.

Components	Concentration	Amount
Capture mAb MT145	0.5 mg/ml	1200 µl
Detection mAb MTG211E, biotin	0.5 mg/ml	50 µl
Streptavidin-HRP	-	500 µl
Polyclonal activator R848	1 mg/ml	100 µl
Lyophilized recombinant human IL-2	-	1 µg

To ensure total recovery of the stated quantity, vials have been overfilled.

Specificity

The kit is based on a matched pair of mAbs enabling specific detection of human IgG2.

Shipping & Storage

- Shipped at ambient temperature.
- Store antibodies and Streptavidin-conjugate at 4-8 °C upon receipt. R848 and IL-2 should be stored at -20 °C.
- Antibodies are supplied in sterile filtered PBS with 0.02% sodium azide. R848 is supplied in PBS with 1% DMSO. Streptavidin-HRP is supplied in PBS with 0.002% Kathon CG.
- The expiry date indicates how long unopened products, stored according to instructions, are recommended for use.

Tips and guidance

Plates

ELISpot plates with a PVDF membrane are recommended. An initial ethanol treatment of the membrane gives optimal binding of capture mAb, ensuring a sensitive assay. It is important that the plate membrane is not allowed to dry after treatment. If drying occurs, repeat protocol steps A2-A3 before adding capture mAb.

Plate washing

Plates can be washed using a multi-channel micropipette. When sterile conditions are not required (protocol steps C1-C5), an ELISA plate washer can be used if the washer head position has been adapted to ELISpot plates. After each wash cycle, empty the plate and tap it gently against absorbent paper.

Serum and cell culture medium

For blocking and cell incubation, we recommend cell culture medium containing 10% fetal calf serum (FCS). The serum should be selected to support cell culture and give low background staining. Homologous serum is not recommended. If required, serum-free medium evaluated for cell culture can be used.

Cells

This assay has been validated with peripheral blood mononuclear cells (PBMCs). For other cell types, the protocol may need adjustment. Freshly prepared or cryopreserved cells can be used. If using frozen cells, leave them diluted in cell culture medium at 37 °C for at least one hour after the thawing procedure. Remove cell debris and count the cells. The assay can be used for analysis of in vivo- as well as in vitro-induced IgG2-secreting cells. For detection of antigen-specific IgG2, the antigen can be coated onto the ELISpot plate or used biotinylated (for the latter, coating should be done with mAb MTG211E, product code: 3852-13).

Stimuli and assay controls

IgG2 secretion can be induced by pre-activation of cells in tubes, with R848 (1 µg/ml) and IL-2* (10 ng/ml), for 3-5 days. After pre-activation, wash the cells to remove secreted IgG2. An assay setup with triplicate wells of 50,000-150,000 cells per well is recommended for total IgG2, and 100,000-500,000 cells per well for antigen-specific IgG2. As negative controls, include wells with unstimulated cells and wells with only medium.

*Reconstitute lyophilized recombinant human IL-2 with 1 ml PBS to obtain 1 µg/ml. Leave for 5 minutes, then vortex. Store in aliquots at -20 °C or below.

Detection antibody

Diluted detection mAb can be filtered (0.2 µm) to reduce the risk of unspecific background.

Analysis

We recommend analyzing plates in an ELISpot reader (e.g., Mabtech ASTOR or Mabtech IRIS).

Please note

- Do not allow the plate membrane to dry during the assay
- Do not touch the membrane with pipette tips
- Avoid getting liquid on the underside of the membrane as it may cause leakage

Protocol

A. Coating (sterile conditions)

1. **Total IgG2:** Dilute the capture mAb (MT145) to 15 µg/ml in sterile PBS.
Antigen-specific IgG2: Dilute the antigen to suitable concentration (1-50 µg/ml) in sterile PBS.
2. Add ethanol (EtOH) to PVDF plates as follows:
 - MSIP plates (Millipore): 20 µl 35% EtOH per well for maximum 1 minute.
 - MAIPSWU plates (Millipore): 50 µl 70% EtOH per well for maximum 2 minutes.
3. Wash the plate 5 times with sterile water, 200 µl per well.
4. Add 100 µl per well of the diluted capture mAb or antigen. Cover the plate and incubate overnight at 4-8 °C.

B. Cell Incubation (sterile conditions)

1. Wash the plate 5 times with sterile PBS, 200 µl per well.
2. Add 200 µl per well of cell culture medium to block and condition the plate. Incubate for at least 30 minutes at room temperature.
3. Empty the plate and add 100 µl cell suspension per well.
4. Place the plate in a 37 °C humidified incubator with 5% CO₂ for 16-24 hours. Avoid evaporation, e.g., by wrapping the plate in aluminum foil. Do not stack or move plates during incubation.

C. Detection

*Use PBS containing 0.5% FCS (PBS-0.5% FCS) for dilution of the detection mAb and Streptavidin-HRP. The PBS should be filtered (0.2 µm) for optimal results. Buffers should **not** contain sodium azide, as it inhibits HRP activity.*

1. Wash the plate 5 times with PBS, 200 µl per well.
2. Add 100 µl per well of the detection mAb (MTG211E-biotin) diluted to 0.5 µg/ml in PBS-0.5% FCS. Incubate for 2 hours at room temperature.
3. Wash as in step C1.
4. Add 100 µl per well of Streptavidin-HRP diluted* in PBS-0.5% FCS. Incubate for 1 hour at room temperature.
*The dilution of Streptavidin-HRP should be adapted to the substrate, e.g., 1:1000 for TMB and 1:100 for AEC.
5. Wash as in step C1.
6. Add 100 µl per well of substrate (e.g., TMB, product code 3651-10) and develop until distinct spots emerge (2-20 minutes).
7. Stop color development by washing thoroughly in deionized water (as ions may cause TMB spots to fade).
8. Leave the plate to dry. The underdrain or tray can be removed for quicker drying. Plates should be completely dry before analysis. For the option to re-analyze later, store plates protected from light at room temperature.

Developed and manufactured by MABTECH AB, Sweden, whose quality management system complies with the standards ISO 9001:2015 & ISO 13485:2016.



The products are for research use only.

MABTECH shall not be liable for the use or handling of the product or for consequential, special, indirect or incidental damages there from.

Mabtech AB (Head Office)
Sweden
Tel: +46 8 716 27 00
mabtech@mabtech.com

Mabtech, Inc.
USA
Tel: +1 513 871-4500
mabtech.usa@mabtech.com