

Human IFN- γ ELISA development kit

Product Code: 3420-1A-20

CONTENTS, development kit for 20 plates:

Vial 1 (green top)

Monoclonal antibody 1-D1K (500 μ l)

Concentration: 1 mg/ml

Vial 2 (yellow top)

Biotinylated monoclonal antibody 7-B6-1 (250 μ l)

Concentration: 1 mg/ml

Vial 3 (white top)

Streptavidin-Alkaline Phosphatase (250 μ l)

Vial 4

Recombinant human IFN- γ standard (1 μ g)

To ensure total recovery of stated quantity, vials have been overfilled.

STORAGE:

Shipped at ambient temperature. On arrival vials should be stored refrigerated at 4-8°C

General

Intended use: For quantitative determination of native and recombinant human IFN- γ in solution, e.g. cell culture supernatant.

Serum/plasma samples: Please note that cytokine determinations in serum/plasma requires the use of ELISA diluent (product code: 3652-D2) for dilution of samples, standard and detection antibody. The diluent prevents false positive read-outs which may be caused by interference of heterophilic antibodies found in plasma and serum. The ELISA diluent has been validated using serum/plasma from normal healthy human blood donors. Please note that heterophilic antibody interference in samples from human subjects with various diseases or other conditions has not been assessed. Please contact Mabtech for further information.

Reagents: Antibodies are supplied in sterile-filtered (0.2 μ m) PBS with sodium azide (0.02%). Streptavidin-ALP is supplied in 0.1 M Tris buffer with 0.002% Kathon CG.


Standard range: 4-400 pg/ml

Limit of detection: 2 pg/ml

Intra-assay variation: < 5%

Standard calibration: 1 ng of supplied standard equals 39 U Gxg01-902-535 NIAID*-standard according to repeated calibrations. Calibration is batch-specific.

*National Institute of Allergy and Infectious Diseases, USA.



Guidelines for Human IFN- γ ELISA

- Day 1**
1. Coat a high protein binding ELISA plate with mAb 1-D1K, diluted to 2 $\mu\text{g}/\text{ml}$ in PBS, pH 7.4, by adding 100 $\mu\text{l}/\text{well}$. Incubate overnight at 4-8°C.
- Day 2**
2. Wash twice with PBS (200 $\mu\text{l}/\text{well}$).
 3. Block plate by adding 200 $\mu\text{l}/\text{well}$ of PBS with 0.05% Tween 20 containing 0.1% BSA (incubation buffer). Incubate for 1 hour at room temperature.
 4. Wash five times with PBS containing 0.05% Tween20
 5. Prepare hIFN- γ standard by reconstituting contents of vial 4 in 1 ml PBS with 1% BSA to a concentration of 1 $\mu\text{g}/\text{ml}$. Leave at room temperature for 15 minutes and then vortex the tube. The stock solution should be used immediately or stored in aliquots at -20°C for future use. We recommend the aliquots not be refrozen after initial use. For the test, prepare dilutions of the stock using the standard range as a guideline.
 6. Add 100 $\mu\text{l}/\text{well}$ of samples or standards diluted in incubation buffer or ELISA diluent for serum/plasma samples and incubate for 2 hours at room temperature.
 7. Wash as in step 4.
 8. Add 100 $\mu\text{l}/\text{well}$ of mAb 7-B6-1-biotin at 1 $\mu\text{g}/\text{ml}$ in incubation buffer or ELISA diluent for serum/plasma samples. Incubate for 1 hour at room temperature.
 9. Wash as in step 4.
 10. Add 100 $\mu\text{l}/\text{well}$ of Streptavidin-ALP diluted 1:1000 in incubation buffer. Incubate for 1 hour at room temperature.
 11. Wash as in step 4.
 12. Add 100 $\mu\text{l}/\text{well}$ of appropriate substrate solution e.g. p-nitrophenyl-phosphate (pNPP), available from Mabtech product code 3652-P10.
 13. Measure the optical density (405 nm for pNPP) in an ELISA reader after suitable developing time.

Developed and manufactured by MABTECH AB, Sweden, whose quality management system complies with the standards ISO 9001:2015 & ISO 13485:2016.



The products are for research use only.

MABTECH shall not be liable for the use or handling of the product or for consequential, special, indirect or incidental damages there from.

Mabtech AB (Head Office)
Sweden
Tel: +46 8 716 27 00
mabtech@mabtech.com

Mabtech, Inc.
USA
Tel: +1 513 871-4500
mabtech.usa@mabtech.com