Mouse IgE
ELISA development kit

Product Code: 3815-1H-6

CONTENTS, development kit for 6 plates:

Vial 1 (green top)
Monoclonal antibody MT56E (300 µl)
Concentration: 0.5 mg/ml

Vial 2 (yellow top)
Biotinylated monoclonal antibody MT44E (80 µl)
Concentration: 0.5 mg/ml

Vial 3 (white top)
Streptavidin-Horseradish Peroxidase (80 µl)

Vial 4
Mouse IgE standard (0.5 µg)

To ensure total recovery of stated quantity, vials have been overfilled.

STORAGE:
Shipped at ambient temperature. On arrival box 1 should be stored refrigerated at 4-8°C and box 2 should be stored frozen at -20°C.

General

Intended use: For quantitative determination of mouse IgE. The ELISA reacts equally with IgE^a (allotype a) and IgE^b (allotype b).

Reagents: Antibodies are supplied in sterile-filtered (0.2 µm) PBS with sodium azide (0.02%). Streptavidin-HRP is supplied in PBS with 1% BSA and 0.002% Kathon CG.

Standard range: 15-1500 pg/ml

Standard calibration: No international standard exists for calibration.



Guidelines for Mouse IgE ELISA

Day 1 1. Coat a high protein binding ELISA plate with mAb MT56E, diluted to 2 µg/ml in PBS, pH 7.4, by adding 100 µl/well. Incubate overnight at 4-8°C.

Day 2 2. Wash twice with PBS (200 µl/well).

3. Block plate by adding 200 µl/well of PBS with 0.05% Tween 20 containing 0.1% BSA (incubation buffer). Incubate for 1 hour at room temperature.

4. Wash five times with PBS containing 0.05% Tween20

5. Prepare mouse IgE standard by reconstituting contents of vial 4 in 1 ml PBS to a concentration of 0.5 µg/ml. Leave at room temperature for 15 minutes and then vortex the tube. The stock solution should be used immediately or stored in aliquots at -20°C for future use. We recommend the aliquots not be refrozen after initial use. For the test, prepare dilutions of the stock using the standard range as a guideline.

6. Add 100 µl/well of samples or standards diluted in incubation buffer and incubate for 2 hours at room temperature.

7. Wash as in step 4.

8. Add 100 µl/well of mAb MT44E-biotin at 0.5 µg/ml in incubation buffer. Incubate for 1 hour at room temperature.

9. Wash as in step 4.

10. Add 100 µl/well of Streptavidin-HRP diluted 1:1000 in incubation buffer. Incubate for 1 hour at room temperature. **Please note that sodium azide used in buffers will inhibit HRP activity.**

11. Wash as in step 4.

12. Add 100 µl/well of appropriate substrate solution e.g. TMB, available from Mabtech product code 3652-F10.

13. Measure the optical density in an ELISA reader after suitable developing time. If required stop the reaction first.

MABTECH

Capture | Detect | Discover

Updated on 2020/03/31

Developed and manufactured by MABTECH AB, Sweden, whose quality management system complies with the standards ISO 9001:2015 & ISO 13485:2016.

CERTIFIED ISO 9001 Quality management systems

CERTIFIED ISO 13485 Medical devices Quality management systems

The products are for research use only.

MABTECH shall not be liable for the use or handling of the product or for consequential, special, indirect or incidental damages there from.

Mabtech AB (Head Office)
Sweden
Tel: +46 8 716 27 00
mabtech@mabtech.com

Mabtech, Inc.
USA
Tel: +1 513 871-4500
mabtech.usa@mabtech.com

www.mabtech.com