# Mouse IgG ELISA development kit

Product Code: 3825-1AD-6

**CONTENTS:** development kit for 6 plates

### Vial 1 (yellow top)

Anti-IgG antibody (150 µl) Concentration: 0.5 mg/ml

## Vial 2 (green top)

ALP-conjugated anti-IgG antibody (80 µl)

#### Vial 3

Lyophilised mouse IgG standard

To ensure total recovery of stated quantity, vials have been overfilled.

#### **STORAGE:**


Shipped at ambient temperature. On arrival box 1 should be stored refrigerated at 4-8°C and box 2 should be stored frozen at -20°C.

## General

**Intended use:** For quantitative determination of native mouse IgG in serum and plasma. May cross-react with immunoglobulins from other species.

**Reagents:** Anti-IgG antibody is supplied in sterile-filtered  $(0.2 \, \mu m)$  PBS with sodium azide (0.02%). ALP-conjugated anti-IgG antibody is supplied in 0.1 M Tris-buffer with 0.002% Kathon CG.

Recommended standard dilution: 0.1-500 ng/ml



# Guidelines for Mouse IgG ELISA

- **Day 1** 1. Coat a high protein binding ELISA plate with anti-IgG antibody, diluted to 1 μg/ml in PBS, pH 7.4, by adding 100 μl/well. Incubate overnight at 4-8°C.
- **Day 2** 2. Wash twice with PBS (200 μl/well).
  - 3. Block plate by adding 200 µl/well of PBS with 0.05% Tween 20 (PBS-Tween) containing 0.1% BSA (incubation buffer\*). Incubate for 1 hour at room temperature.
  - 4. Wash five times with PBS-Tween.
  - 5. Prepare mouse IgG standard by reconstituting contents of vial 3 in 500 μl PBS to make up a stock solution of 50 μg/ml. The stock solution should be used immediately or stored in aliquots at -20°C for future use. We recommend the aliquots not be refrozen after initial use. For the test, prepare dilutions of the stock using the standard range as a guideline.
  - 6. Add 100 µl/well of samples or standards diluted in incubation buffer and incubate for 2 hours at room temperature.
  - 7. Wash as in step 4.
  - 8. Add 100 µl/well of anti-IgG-ALP diluted 1:1000 in incubation buffer. Incubate for 1 hour at room temperature.
  - 9. Wash as in step 4.
  - 10. Add 100 μl/well of appropriate substrate solution e.g. p-nitrophenyl-phosphate (pNPP).
  - 11. Measure the optical density (405 nm for pNPP) in an ELISA reader after suitable developing time.
    - \* The same buffer is used for blocking and for dilution.



Developed and manufactured by MABTECH AB, Sweden, whose quality management system complies with the standards ISO 9001:2015 & ISO 13485:2016.





#### The products are for research use only.

MABTECH shall not be liable for the use or handling of the product or for consequential, special, indirect or incidental damages there from.

Mabtech AB (Head Office)

Sweden Tel: +46 8 716 27 00 mabtech@mabtech.com Mabtech, Inc.

USA Tel: +1 513 871-4500 mabtech.usa@mabtech.com