

Monkey TNF- α FluoroSpot^{BASIC} (550)

PRODUCT CODE: FS1-28-550

| | |
|-------------------------------|---|
| Capture antibody: | mAb MT21A8, 0.5 mg/ml (1200 μ l) |
| Detection antibody: | mAb MT15B15, biotinylated, 0.5 mg/ml (180 μ l) |
| Fluorophore conjugate: | SA-550 (240 μ l) |
| Fluorescence enhancer: | Ready-to-use Fluorescence enhancer-II (25 ml) |
| Supplied in: | Monoclonal antibodies are supplied in sterile filtered (0.2 μ m) PBS with 0.02% sodium azide. Streptavidin-conjugate SA-550 and the Fluorescence enhancer contain 0.002% Kathon CG. |
| Quantity: | Reagents for 4 plates |
| Storage: | Shipped at ambient temperature. On arrival, store product refrigerated at 4-8°C. |
| Specificity: | The monoclonal antibodies react with human TNF- α and cross-react with TNF- α from non-human primates. Please visit www.mabtech.com/NHP for information on reactivity with monkey species. |
| Applications: | For enumeration of TNF- α secreting cells using FluoroSpot. Please find the recommended assay conditions below. |
| Analysis: | A FluoroSpot reader equipped with a filter for LED550, Cy3 (excitation 550 nm/emission 570 nm) is recommended for analysis. |
| Protocols: | General protocol for the FluoroSpot assay is provided in the package and is also available at www.mabtech.com/fluorospot-protocol . It is recommended to rest cryopreserved cells for at least one hour to allow removal of cell debris. To decrease background TNF- α spots, a subsequent resting overnight is recommended after removal of cell debris. |
| Assay conditions: | Reagents should not be diluted in advance |

| | | |
|--------------------------------------|---|-------------------------|
| Capture antibody | MT21A8 | Dilute to 15 μ g/ml |
| Detection antibody | MT15B15, biotinylated | Dilute to 2 μ g/ml |
| Fluorophore conjugate | SA-550 | Dilute 1:200 |
| Fluorescence enhancer | Fluorescence enhancer-II | 50 μ l/well |
| Cell incubation time | 12-48 hours in the plate | |
| Recommended positive controls | PHA (1-10 μ g/ml) or LPS (1-10 ng/ml) | |



Note; for research use only.

Mabtech shall not be liable for the use or handling of the product or for consequential, special, indirect or incidental damages therefrom.



Developed and manufactured by MABTECH AB, Sweden, whose quality management system complies with the above standards.

Mabtech AB (Head Office)
Sweden
Tel +46 8 716 27 00
mabtech@mabtech.com
Augustendalstorget 9
SE - 131 52 Nacka Strand

Mabtech Inc.
USA
Tel: +1 513 871-4500
mabtech.usa@mabtech.com

Mabtech AB, Büro Deutschland
Germany
Tel: +49 40 4135 7935
mabtech.de@mabtech.com

Mabtech AB, Bureau de liaison
France
Tel: +33 (0)4 92 38 80 70
mabtech.fr@mabtech.com

Mabtech Australia Pty Ltd
Australia
Tel: +61 3 9470 4704
mabtech.au@mabtech.com